

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 8528 (1988): Pliers, Anterior Band Forming, Double Beak, Dental [MHD 8: Dentistry]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE



Indian Standard

(Reaffirmed 1993)

SPECIFICATION FOR PLIERS, ANTERIOR BAND FORMING, DOUBLE BEAK, DENTAL


(First Revision)

1. Scope — This standard prescribes requirements for anterior band forming double beak pliers used in dentistry.

2. Material — As specified in IS : 12507-1988 'General requirements for dental pliers'.

2.1 Spring — Stainless steel wire conforming to IS : 4454 (Part 4)-1974 'Specification for steel wires for cold formed springs: Part 4 Stainless spring steel wire for normal corrosion resistance (first revision)'.

3. Shape and Dimensions — As shown in Fig. 1.



All dimensions in millimetres.

FIG. 1 PLIERS, ANTERIOR BAND FORMING, DOUBLE BEAK, DENTAL

Adopted 24 October 1988

© March 1989, BIS

Gr 1

IS : 8528 – 1988

3.1 The studs shall be push fit and their size shall be adequate to enable the maximum opening of the beaks of the pliers and to avoid the slipping of the spring.

4. Hardness — 370 to 410 HV.

5. Other Requirements — As laid down in IS : 12507-1988.

EXPLANATORY NOTE

This standard was first published in 1977. The revision has been brought out to take into account the provisions of the Indian Standard on general requirements for dental pliers, IS : 12507-1988.